



त्रिभुवन विश्वविद्यालय

शिक्षण अस्पताल

तार-टुथमेड
महाराजगञ्ज
काठमाडौं, नेपाल ।

पत्र संख्या :-



मिति :

सूचना !

२०८१।०१।१३

यस अस्पतालको प्लास्टिक सर्जरी तथा वर्न विभागको लागि आवश्यक एक थान **Battery Operated (Electric) SAW for Bone** खरिद कार्य गर्नको लागि प्रतिस्पर्धात्मक दररेटमा शिलबन्दी कोटेशन आह्वान प्रयोजनार्थ यो सूचना प्रकाशित भएको मितिले सात(७) दिनभित्र ईच्छुक इजाजत प्राप्त सप्लायर्सहरूले आवश्यक सम्पूर्ण कागजातका साथै दररेट विवरण सहित शिलबन्दी कोटेशन अस्पतालको सामान्य प्रशासन शाखा "क" मा कार्यालय समय भित्र पेश गर्न सूचित गरिन्छ । यसैसाथ संलग्न प्राविधिक विवरण बमोजिमको शिलबन्दी कोटेशनमा आवश्यक सम्पूर्ण कागजपत्रहरू अनिवार्य रूपमा संलग्न हुनुपर्नेछ । थप जानकारीको लागि अस्पतालको सूचना पाटी र अस्पतालको सामान्य प्रशासन शाखा "क" मा सम्पर्क गर्न सकिनेछ ।

...
यादव प्रसाद पोखरेल

यादव प्रसाद पोखरेल

प्रमुख

सामान्य प्रशासन क

Dinesh Kafle



1. Price Schedule for Machine

1	2	3	4	5	6	7	8
Item	Description	Unit	Quantity	Unit price (Site Delivery)	Total price in figure (cols. 4 x 5)	Total price in words	Remarks
1	Battery Operated (Electric) SAW for Bone	Set	One				
Total Amount							
Add 13% Value Added Tax							
Total Including VAT							

Total Price (in words)

Signature and Stamp of Bidder _____

Note: In case of discrepancy between unit price and total, the unit price shall prevail

2. Schedule of Requirements

The delivery schedule expressed as days/weeks/months stipulates hereafter a delivery date which is the date of delivery to the final destination where the Goods is required to be delivered.

No.	Description	Quantity	Place of Delivery	Delivery schedule days/weeks/months from date of Purchase Order
1	Battery Operated (Electric) SAW for Bone	One	TUTH, Maharajgunj, Ktm.	Seven Days

3. Technical Specifications

Technical specification of battery operate saw for the plastic surgery		
Name of the bidder		
Model no.		
Country of origin		
S.no	Specifications	References
1.	Utilizes a rechargeable and lightweight battery for enhanced portability and maneuverability during plastic surgery procedures.	
2.	Features a precision motor and blade system designed for delicate and controlled cuts in soft tissues such as skin and fat.	
3.	Adjustable speed settings to accommodate various plastic surgery procedures.	
4.	Speed 0-20000 OSC/Min.	
5.	Constructed with materials suitable for standard sterilization methods, including autoclaving, to maintain a sterile surgical environment.	
6.	Easily disassembled for thorough cleaning and sterilization between procedures.	
7.	Compatible with a range of specialized blades designed for specific plastic surgery applications, allowing for versatility in procedures.	
8.	Quick and secure blade attachment mechanism for efficient blade changes during surgery.	
9.	Integrated safety features to prevent accidental injuries to both the surgeon and the patient.	
10.	Mechanisms in place to minimize heat generation and reduce the risk of thermal injury to tissues.	
11.	Ergonomic design to ensure comfort during prolonged use and to facilitate precise control for intricate plastic surgery procedures. Lightweight and balanced for ease of handling.	
12.	Low noise less than 60 dB and vibration levels to enhance patient comfort and minimize potential disruption to the surgical team.	
13.	Designed to meet or exceed industry standards for noise and vibration in medical devices.	
14.	Utilizes biocompatible materials to minimize the risk of tissue irritation or allergic reactions during and after surgery. Materials chosen for easy cleaning and compatibility with sterilization processes.	
15.	Materials chosen for easy cleaning and compatibility with sterilization processes.	
Terms and Conditions		
16.	Compliance with relevant medical device regulations and standards specific to plastic surgery applications.	
17.	Compliance with relevant medical device regulations and standards, such as ISO and FDA requirements.	
18.	Comprehensive documentation for regulatory submissions, including design specifications, testing results, and quality control measures.	
19.	Should provide 1 years of full Warranty service and parts.	
20.	Should provide service and user manual Hard copy and soft copy.	
21.	Suppliers should provide valid authorization letter.	
22.	Suppliers should provide Parts list with price.	