



त्रिभुवन विश्वविद्यालय

शिक्षण अस्पताल

तार-दुधमेड
महाराजगञ्ज
काठमाडौं, नेपाल ।

पत्र संख्या :-



मिति :

सूचना !

२०८१।०१।१८

यस अस्पतालको IVF विभागमा Zeiss Company को *STEMI 508 Model* को Microscope को लागि आवश्यक एक थान **Microscopic Camera** खरिद कार्य गर्नको लागि प्रतिस्पर्धात्मक दररेटमा शिलबन्दी कोटेशन आह्वान प्रयोजनार्थ यो सूचना प्रकाशित भएको मितिले सात(७) दिनभित्र ईच्छुक इजाजत प्राप्त सप्लायर्सहरुले आवश्यक सम्पूर्ण कागजातका साथै दररेट विवरण सहित शिलबन्दी कोटेशन अस्पतालको सामान्य प्रशासन शाखा "क" मा कार्यालय समय भित्र पेश गर्न सूचित गरिन्छ । यसैसाथ संलग्न प्राविधिक विवरण बमोजिमको शिलबन्दी कोटेशनमा आवश्यक सम्पूर्ण कागजपत्रहरु अनिवार्य रुपमा संलग्न हुनुपर्नेछ । थप जानकारीको लागि अस्पतालको सामान्य प्रशासन शाखा "क" मा सम्पर्क गर्न सकिनेछ ।



यादव प्रसाद पोखरेल
प्रमुख
सामान्य प्रशासन 'क'



1. Price Schedule for Microscopic Camera

1	2	3	4	5	6	7	8
Item	Description	Unit	Quantity	Unit price (Site Delivery)	Total price in figure (cols. 4 x 5)	Total price in words	Remarks
1	Microscopic Camera	Set	One				
Total Amount							
Add 13% Value Added Tax							
Total Including VAT							

Total Price (in words)

Signature and Stamp of Bidder _____

Note: In case of discrepancy between unit price and total, the unit price shall prevail

2. Schedule of Requirements

The delivery schedule expressed as days/weeks/months stipulates hereafter a delivery date which is the date of delivery to the final destination where the Goods is required to be delivered.

No.	Description	Quantity	Place of Delivery	Delivery schedule days/weeks/months from date of Purchase Order
1	Microscopic Camera	One Set	TUTH, Maharajgunj, Ktm.	Seven Days

3. Technical Specifications

Technical specification of Microscopic camera		
Name of Bidder:		Remarks
Brand:		
Manufacturing Company:		
Model :		
Country of Origin:		
S.N	Specifications	
1.	Microscopy color camera at least 8 mega pixels.	
2.	Should have pixel size of at least 1.85 μ m*1.85 μ m.	
3.	Should have exposure range minimum 0.06ms to 1s.	
4.	Should have Bit depth of approximately 3*8bit/pixel.	
5.	Should have passive cooling system.	
6.	Power supply should be compatible to international sockets available.	
7.	System should have following interfaces; a.HDMI b.USB3.0 c.Type C d. Ethernet e.Micro D	
8.	Should have frame rate of not less than 30fps.	
9.	Should have facility of image snap with single button.	
10.	Should have Wifi compatibility .	
11.	Should have automatic adjustment of parameters and scaling information.	
Terms and conditions		
12.	Camera should have compatibility with existing microscope at hospital.	
13.	Should provide 1 years of full Warranty service and parts.	
14.	Suppliers should provide valid authorization letter.	
15.	User Manual: Clear instructions on application and safety.	