



त्रिभुवन विश्वविद्यालय

शिक्षण अस्पताल

तार-दुधमेड
महाराजगञ्ज
काठमाडौं, नेपाल।

पत्र संख्या :-

मिति :

सूचना

२०८१।०२।०८

यस अस्पतालको लागि CO2 Insufflator दूई थान आवश्यक भएको हुनाले उक्त सामग्री आपूर्ति गर्न इच्छुक इजाजत प्राप्त सप्लायर्स/कम्पनी/फर्महरुले यो सूचना प्रकाशित मितिले सात दिन भित्र आवश्यक सम्पूर्ण कागजातका साथै दररेट विवरण सहित शिलबन्दी कोटेशन अस्पतालको सामान्य प्रशासन शाखा "क" मा पेश गर्नुहुन सूचित गरिन्छ।

यादव प्रसाद पोखरेल
प्रमुख
सामान्य प्रशासन शाखा क

Dinesh Kafle



1. Price Schedule for Machine

1	2	3	4	5	6	7	8
Item	Description	Unit	Quantity	Unit price (Site Delivery)	Total price in figure (cols. 4 x 5)	Total price in words	Remarks
1	CO2 Insufflator	Set	Two				
Total Amount							
Add 13% Value Added Tax							
Total Including VAT							

Total Price (in words)

Signature and Stamp of Bidder _____

Note: In case of discrepancy between unit price and total, the unit price shall prevail

2. Schedule of Requirements

The delivery schedule expressed as days/weeks/months stipulates hereafter a delivery date which is the date of delivery to the final destination where the Goods is required to be delivered.

No.	Description	Quantity	Place of Delivery	Delivery schedule days/weeks/months from date of Purchase Order
1	Battery Operated (Electric) SAW for Bone	One	TUTH, Maharajgunj, Ktm.	Seven Days

3. Technical Specifications

S.N	Purchaser's Technical specifications	YES/NO or Any Deviation
1.	Name of the Bidder:	
2.	Manufacturer:	
3.	Model:	
4.	Country Of Origin/Made in :	
5.	Should be Microprocessor based system with flow rate setting of at least 1 to 30 litre/min.	
6.	System should have insufflation pressure range setting from 5 to 30 mmHg or better	
7.	System should have digital display for actual gas flow rate, set gas flow rate, Actual gas pressure, set gas pressure, temperature and LED indicator for start/stop operation and for any fault in machine.	
8.	System should have inbuilt gas heater.	
9.	System should have Co2 gas status indicator for indicating the flow of gas from cylinder.	
10.	System should be compatible and should adapt the existing operation theatre systems.	
11.	Appropriate standard accessories including following items(for each machine):- a) Silicone insufflation tube 1 set b) Power cable to connect mains ac 1 set c) Universal wrench for cylinder 1 set d) Set of appropriate gas connector with high pressure tubing, regulator and flow meter to connect co2 cylinder to the insufflator 1set	
12.	Operating voltage should be 220V+-10%,50Hz AC.	
	Terms and Conditions	
1.	The unit shall have European CE or USFDA certified. Bidder must submit the certificate with bidding document	
2.	Must submit original brochure.	
3.	Should submit authorization from manufacturer	
4.	Bidder must provide Comprehensive warranty for 1 year.	
5.	All the necessary work for the complete installation of unit should be done by bidder.	
6.	User training should be provided for OT staff.	