



त्रिभुवन विश्वविद्यालय

शिक्षण अस्पताल

तार-दुधमेड
महाराजगञ्ज
काठमाडौं, नेपाल।

पत्र संख्या :-



मिति :

सूचना !

२०८१।०२।१०

यस अस्पतालको लागि आवश्यक दुई थान **Patient Warmer Device (Blower)** खरिद कार्य गर्नको लागि प्रतिस्पर्धात्मक दररेटमा शिलबन्दी कोटेशन आह्वान प्रयोजनार्थ यो सूचना प्रकाशित भएको मितिले सात(७) दिनभित्र ईच्छुक इजाजत प्राप्त सप्लायर्सहरूले आवश्यक सम्पूर्ण कागजातका साथै दररेट विवरण सहित शिलबन्दी कोटेशन अस्पतालको सामान्य प्रशासन शाखा "क" मा कार्यालय समय भित्र पेश गर्न सूचित गरिन्छ। यसैसाथ संलग्न प्राविधिक विवरण बमोजिमको शिलबन्दी कोटेशनमा आवश्यक सम्पूर्ण कागजपत्रहरू अनिवार्य रूपमा संलग्न हुनुपर्नेछ। थप जानकारीको लागि अस्पतालको सामान्य प्रशासन शाखा "क" मा सम्पर्क गर्न सकिनेछ।



यादव प्रसाद पोखरेल

प्रमुख

सामान्य प्रशासन 'क'

Dinesh Kafle



1. Price Schedule for Machine

1	2	3	4	5	6	7	8
Item	Description	Unit	Quantity	Unit price (Site Delivery)	Total price in figure (cols. 4 x 5)	Total price in words	Remarks
1	Patient Warmer Device (Blower)	Set	Two				
Total Amount							
Add 13% Value Added Tax							
Total Including VAT							

Total Price (in words)

Signature and Stamp of Bidder _____

Note: In case of discrepancy between unit price and total, the unit price shall prevail

2. Schedule of Requirements

The delivery schedule expressed as days/weeks/months stipulates hereafter a delivery date which is the date of delivery to the final destination where the Goods is required to be delivered.

No.	Description	Quantity	Place of Delivery	Delivery schedule days/weeks/months from date of Purchase Order
1	Patient Warmer Device (Blower)	Two	TUTH, Maharajgunj, Ktm.	Seven Days

3. Technical Specifications of BODY WARMING DEVICE with mobile trolley

S.No.	Technical Specifications
	Name of the bidder:
	Manufacturer:
	Made in:
	Model No:
1	The equipment should work on the principles of forced air/filtered air warming.
2	The equipment should have 0.2 micron filter or better for air filtration.
3	The equipment should operate in at least three ranges of temperatures: High: 43 +/- 3 deg C, Med: 38 +/- 3 deg C, low: 32 +/- 3 deg C.
4	The equipment should operate on 100-240 VAC, 50-60 Hz
5	System should have same company manufactured trolley/stand with wheel for easy movement.
6	The unit should be supplied with all standard accessories including warming hose, trolley/stand, power cord etc.
Terms and Conditions	
7	The unit should have European CE and US FDA certificate. The supplier must submit the Certificate.
8	The supplier must submit the original brochure or e-copy.
9	Should have 1 years complete parts & service warranty.